

FICHA TÉCNICA

CALIBRADOR		
REFERENCIA	80256	
FABRICANTE / RESPONSABLE	MARK ONE SURGICAL Ugoki Road, Adalat Garah, Sialkot 51330, Pakistan	
DESCRIPCIÓN	Calibrador decimal.	
ESPECIFICACIÓN	Puntas finas para hacer medidas hasta 10 cm Medidas (cm): 15 cm de largo x 1 punta	
COMPOSICIÓN	Acero inoxidable.	
VIDA ÚTIL	La vida útil del instrumento depende de la frecuencia de uso, el cuidado del usuario y los métodos de reprocesamiento adecuados.	
PRESENTACIÓN	Contiene una unidad.	
FINALIDAD	Instrumentos para coronas.	
ESTERILIZACIÓN	Todos los instrumentos se deben examinar, limpiar y esterilizar antes de cada uso. Esterilizar a 177 °C en autoclave durante al menos 4 minutos, seguido de un secado adecuado.	
ALMACENAMIENTO Y TRANSPORTE	No requiere condiciones especiales de almacenamiento y transporte.	

TECHNICAL DATA SHEET

CALIBRATOR		
REFERENCE	80256	
MANUFACTURER	MARK ONE SURGICAL Ugoki Road, Adalat Garah, Sialkot 51330, Pakistan	
DESCRIPTION	Decimal calibrator.	
SPECIFICATION	Thin tips to make measurements up to 10 cm Measurements (cm): 15 cm long x 1 tip.	
COMPOSITION	Stainless steel.	
SHELF LIFE	Instrument lifetime depends on frequency of use, user care and proper reprocessing methods.	
PACKAGE	1 unit.	
INTENDED USE	Instruments for crowns.	
STERILIZATION	All instruments should be examined, cleaned and sterilized before each use. Sterilize at 177 oC in autoclave for at least 4 minutes, followed by adequate drying.	
STORAGE AND TRANSPORT	It does not require special storage and transportation conditions.	